

High Performance Liquid Chromatography


1. Analysis Information

Product Name:	Q-VD-OPh
Operator:	David
Injection Date:	3/09/2014 15:30 PM
Batch No.:	2

2. HPLC Condition

Column:	xBridge C18,4.6*150mm,3.5 μ m										
Solvent A:	0.1%Trifluoroacetic in 100% Acetonitrile										
Solvent B:	0.1%Trifluoroacetic in 100% Water										
Gradient:	<table border="1"> <thead> <tr> <th>Time</th> <th>A</th> <th>B</th> </tr> </thead> <tbody> <tr> <td>0.0min</td> <td>30%</td> <td>70%</td> </tr> <tr> <td>15.0min</td> <td>70%</td> <td>30%</td> </tr> </tbody> </table>		Time	A	B	0.0min	30%	70%	15.0min	70%	30%
Time	A	B									
0.0min	30%	70%									
15.0min	70%	30%									
Flow rate:	0.6ml/min										
Wavelength:	254nm										
Volume:	10 μ l										

3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	5.941	1030074	3.87
2	8.074	13338587	50.05
3	8.306	12050175	45.22
4	9.820	231309	0.87