

High Performance Liquid Chromatography

1. Analysis Information

Product Name:	Z-VAD-FMK
Operator:	David
Injection Date:	2015-8-2 9:37:42
Batch No.:	9


2. HPLC Condition

Column:	Athena C18, 3 μ m, 2.1mm \times 100 mm
Solvent A:	0.1% H ₃ PO ₄ in 100% Acetonitrile
Solvent B:	0.1% H ₃ PO ₄ in 100% Water

Gradient:			
	Time	A	B
	0.0min	0%	100%
	2.0min	0%	100%
	4.0min	95%	5%
	10.0min	0%	100%
	10.0min	Stop	Stop

Flow rate:	0.4ml/min
Wavelength:	215nm
Volume:	10 μ l

3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	2.726	80.31229	0.4793
2	2.889	27.31856	0.1630
3	3.540	1.63436e4	97.5347
4	4.110	212.24371	1.2666
5	4.636	93.22586	0.5563