

## Mass Spectrometry

### Analysis Information

**Product Name:** Alarelin Acetate  
**Operator:** David  
**Injection Date:** 7/15/2020 10:33:34 AM  
**Inj. Volume:** 1  
**Batch No.:** 4

### Acquisition Parameter


**Probe:** ESI  
**Gas Temp:** 350°C  
**Gas Flow:** 11L/min  
**Nebulizer:** 50psi  
**VCap:** 4000v  
**Fragmentor:** 70v  
**Reference:** m/z 121.050873, 922.009798  
**T.Flow:** 0.3ml/min  
**B.conc:** 20%H<sub>2</sub>O/80%ACN

## Result

NT-S20200715-006

20200715-A3151-4 11 (0.182)

1: Scan ES+  
6.20e6



Molecular Weight: 1287.42

m/z([M-2C<sub>2</sub>H<sub>4</sub>O<sub>2</sub>]-) Base Peak: 1168.32

## Caution

**FOR RESEARCH PURPOSES ONLY.**

**NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.**

**APEX BIO Technology**

**www.apexbt.com**

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com