



High Performance Liquid Chromatography


1. Analysis Information

Product Name:	AM095
Operator:	David
Injection Date:	10/11/2015 8:55:06 AM
Batch No.:	1

2. HPLC Condition

Column:	Athena C18, 3 μ m, 2.1mm \times 100 mm																			
Solvent A:	0.1% H3PO4 in 100% Acetonitrile																			
Solvent B:	0.1% H3PO4 in 100% Water																			
Gradient:	<table border="1"> <thead> <tr> <th>Time</th> <th>A</th> <th>B</th> </tr> </thead> <tbody> <tr> <td>0.0min</td> <td>0%</td> <td>100%</td> </tr> <tr> <td>5.0min</td> <td>0%</td> <td>100%</td> </tr> <tr> <td>28.0min</td> <td>95%</td> <td>5%</td> </tr> <tr> <td>40.0min</td> <td>0%</td> <td>100%</td> </tr> <tr> <td>45.0min</td> <td>Stop</td> <td>Stop</td> </tr> </tbody> </table>		Time	A	B	0.0min	0%	100%	5.0min	0%	100%	28.0min	95%	5%	40.0min	0%	100%	45.0min	Stop	Stop
Time	A	B																		
0.0min	0%	100%																		
5.0min	0%	100%																		
28.0min	95%	5%																		
40.0min	0%	100%																		
45.0min	Stop	Stop																		
Flow rate:	0.4ml/min																			
Wavelength:	220nm																			
Volume:	10 μ l																			

3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	28.288	40754637	99.669
2	32.893	135282	0.331