



APEX BIO
Achieve Perfection, Explore the Unknown

High Performance Liquid Chromatography


1. Analysis Information

Product Name:	AUY922 (NVP-AUY922)
Operator:	David
Injection Date:	12/2/2013 1:30:55 PM
Batch No.:	1

2. HPLC Condition

Column:	Venusil ASB C18,4.6*250mm,5µm,300A																
Solvent A:	0.1% Trifluoroacetic in 100% Acetonitrile																
Solvent B:	0.1% Trifluoroacetic in 100% Water																
Gradient:	<table border="1"> <thead> <tr> <th>Time</th> <th>A</th> <th>B</th> </tr> </thead> <tbody> <tr> <td>0.0min</td> <td>95%</td> <td>5%</td> </tr> <tr> <td>3.0min</td> <td>95%</td> <td>5%</td> </tr> <tr> <td>6.0min</td> <td>95%</td> <td>5%</td> </tr> <tr> <td>6.0min</td> <td>Stop</td> <td>Stop</td> </tr> </tbody> </table>		Time	A	B	0.0min	95%	5%	3.0min	95%	5%	6.0min	95%	5%	6.0min	Stop	Stop
Time	A	B															
0.0min	95%	5%															
3.0min	95%	5%															
6.0min	95%	5%															
6.0min	Stop	Stop															
Flow rate:	1.0ml/min																
Wavelength:	254nm																
Volume:	10µl																

3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	3.307	3244.50269	97.8667
2	3.431	35.21147	1.0621
3	4.042	35.51323	1.0712