



High Performance Liquid Chromatography


1. Analysis Information

Product Name:	BX795
Operator:	David
Injection Date:	3/4/2014 6:21:06 PM
Batch No.:	1

2. HPLC Condition

Column:	Athena C18 (3 μ m, 2.1 mm \times 100 mm)																						
Solvent A:	0.1% H3PO4 in 100% Acetonitrile																						
Solvent B:	0.1% H3PO4 in 100% Water																						
Gradient:	<table border="1"><thead><tr><th>Time</th><th>A</th><th>B</th></tr></thead><tbody><tr><td>0.0min</td><td>10%</td><td>90%</td></tr><tr><td>3.5min</td><td>95%</td><td>5%</td></tr><tr><td>8.0min</td><td>95%</td><td>5%</td></tr><tr><td>9.0min</td><td>10%</td><td>90%</td></tr><tr><td>10.0min</td><td>10%</td><td>90%</td></tr><tr><td>10.0min</td><td>Stop</td><td>Stop</td></tr></tbody></table>		Time	A	B	0.0min	10%	90%	3.5min	95%	5%	8.0min	95%	5%	9.0min	10%	90%	10.0min	10%	90%	10.0min	Stop	Stop
Time	A	B																					
0.0min	10%	90%																					
3.5min	95%	5%																					
8.0min	95%	5%																					
9.0min	10%	90%																					
10.0min	10%	90%																					
10.0min	Stop	Stop																					
Flow rate:	0.4ml/min																						
Wavelength:	220nm																						
Volume:	10 μ l																						

3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	3.450	445.81631	3.1708
2	3.536	1.32563e4	94.2818
3	3.746	128.94518	0.9171
4	3.788	130.98698	0.9316
5	3.883	98.24700	0.6988