

High Performance Liquid Chromatography

Analysis Information

Product Name: OAC2
Operator: David
Injection Date: 5/12/2020 3:28:44 PM
Batch No.: 2

HPLC Condition


Column : ZORBAX Eclipse Pluse C18 3.5µm 4.6mmx100mm
Solvent A: H2O+0.1%TFA
Solvent B: ACN+0.1%TFA

Gradient:

Time	A	B
0.0min	80%	20%
2.0min	80%	20%
7.0min	5%	95%
10.00min	0%	100%
12.00min	0%	100%
15.00min	Stop	Stop

Flow rate: 1mL/min
Wavelength: 330nm
Volume: 5µl

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	6.283	921.15070	97.7432
2	6.551	21.26874	2.2568

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com