



# High Performance Liquid Chromatography

## 1. Analysis Information

Product Name: AZD-9291  
Operator: David  
Injection Date: 6/19/2014 9:01:38 AM  
Batch No.: 2


## 2. HPLC Condition

Column: Athena C18, 3  $\mu$ m, 2.1mm  $\times$  100 mm  
Solvent A: 0.1% H<sub>3</sub>PO<sub>4</sub> in 100% Acetonitrile  
Solvent B: 0.1% H<sub>3</sub>PO<sub>4</sub> in 100% Water

Gradient:	Time	A	B
	0.0min	10%	90%
	3.5min	95%	5%
	8.0min	95%	5%
	9.0min	10%	90%
	10.0min	10%	90%
	10.0min	Stop	Stop

Flow rate: 0.4ml/min  
Wavelength: 254nm  
Volume: 10 $\mu$ l

## 3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	3.232	1988.79333	98.8199
2	3.501	23.75051	1.1801