

High Performance Liquid Chromatography

Analysis Information

Product Name: SCH772984 HCl
Operator: David
Injection Date: 4/13/2021 1:58:56 AM
Batch No.: 2


HPLC Condition

Column : ZORBAX Eclipse Pluse C18 3.5 μ m 4.6mmx100mm
Solvent A: H2O+0.1%TFA
Solvent B: ACN+0.1%TFA
Gradient:

Time	A	B
0.0min	80%	20%
2.0min	80%	20%
16.0min	20%	80%
17.0min	0%	100%
20.0min	0%	100%
23.0min	Stop	Stop

Flow rate: 1ml/min
Wavelength: 254nm
Volume: 5 μ l

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	2.123	29.17307	0.2397
2	5.211	1.18708e4	97.5266
3	5.970	22.55413	0.1853
4	6.254	27.10098	0.2227
5	6.501	32.26836	0.2651
6	6.623	18.18950	0.1494
7	7.114	49.64825	0.4079
8	7.226	50.40173	0.4141
9	7.477	71.71783	0.5892

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com

