

High Performance Liquid Chromatography

Analysis Information

Product Name: GSK1016790A
Operator: David
Injection Date: 10/26/2021 6:36:13 PM
Batch No.: 2


HPLC Condition

Column : Athena C18, 3 μ m, 2.1mm \times 100 mm
Solvent A: ACN+0.1%TFA
Solvent B: H2O+0.1%TFA
Gradient:

Time	A	B
0.0min	10%	90%
5.0min	95%	5%
9.0min	95%	5%
9.0min	Stop	Stop

Flow rate: 0.46ml/min
Wavelength: 210nm
Volume: 3 μ l

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	4.211	5.27645	0.0640
2	5.711	7.72771	0.0937
3	5.874	8203.93066	99.4551
4	6.519	6.20801	0.0753
5	6.714	4.69812	0.0570
6	7.361	15.65034	0.1897
7	7.426	5.38750	0.0653

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com

