

Product Information

Caspase-8 Fluorometric Assay Kit

I. Kit Contents:

| Component | K2012-25 25 assays | K2012-100 100 assays | K2012-200 200 assays | K2012-400 400 assays | Part Number |
|--------------------|-----------------------|-------------------------|-------------------------|-------------------------|-------------|
| Cell Lysis Buffer | 25 ml | 100 ml | 100 ml | 100 ml | K2012-C-1 |
| 2X Reaction Buffer | 2 ml | 4 x 2 ml | 16 ml | 32 ml | K2012-C-2 |
| IETD-AFC (1 mM) | 125 μ l | 500 μ l | 2 x 0.5 ml | 2 x 1 ml | K2012-C-3 |
| DTT (1 M) | 100 μ l | 400 μ l | 400 μ l | 400 μ l | K2012-C-4 |

II. Introduction:

Cysteine-dependent aspartate-directed proteases (Caspases) are a family of cysteine proteases that play important roles in apoptosis, necrosis, and inflammation. Sequential activation of caspases plays an important role in cell apoptosis. Caspase-8 is a caspase protein that cleaves and activates Caspase-3 and is involved in the programmed cell death induced by Fas and different apoptotic stimuli. Caspase-8 was detected in the insoluble fraction of the affected brain region in Huntington disease patients, which suggested it plays an important role in neurodegenerative diseases.

Caspase-8 Fluorometric Assay Kit provides a convenient and simple way for detecting the IETD-dependent caspase activity. IETD-AFC (AFC: 7-amino-4-trifluoromethyl coumarin) emits blue light ($\lambda_{max} = 400$ nm); while cleavage of IETD-AFC by Caspase-8 or related caspases, free AFC emits a yellow-green fluorescence ($\lambda_{max} = 505$ nm), which can be quantified by using a fluorescence microtiter plate reader or a fluorometer. Comparison of the fluorescence of AFC from an apoptotic sample with an uninduced control determines the fold increase in Caspase-8 activity.

III. Caspase-8 Assay Protocol:

A. General Considerations

Aliquot enough 2X Reaction Buffer for the number of assays to be performed. Add DTT to the 2X Reaction Buffer immediately before use (10 mM final concentration: add 10 μ l of 1.0 M DTT stock per 1 ml of 2X Reaction Buffer).

After thawing, store the Cell Lysis Buffer and 2X Reaction Buffer at 4 °C.

Protect IETD-AFC from light.

B. Assay Procedure

1. Induce apoptosis in cells by desired method. Concurrently incubate a control culture without induction.
2. Count cells and pellet $1-5 \times 10^6$ cells or use 50-200 μ g cell lysates if protein concentration has been measured.
3. Resuspend cells in 50 μ l of chilled Cell Lysis Buffer. Incubate cells on ice for 10 minutes.
4. Add 50 μ l of 2X Reaction Buffer (containing 10 mM DTT) to each sample. Add 5 μ l of the 1 mM IETD-AFC substrate (50 μ M final concentration). Incubate at 37 °C for 1-2 hours.
5. Read samples in a fluorometer equipped with a 400-nm excitation filter and 505-nm emission filter. For a plate-reading set-up, transfer the samples to a 96-well plate. You may also perform the entire assay directly in a 96-well plate.

Fold-increase in FLICE activity can be determined by comparing these results with the level of the uninduced control.

IV. Storage and Stability:

Store kit at -20 (Store Cell Lysis Buffer and 2X Reaction Buffer at 4°C after opening). All reagents are stable for 6 months under proper storage conditions.

General Troubleshooting Guide for Caspase Colorimetric and Fluorometric Kits:

| Problems | Cause | Solution |
|-------------------------------|--|---|
| Assay not working | <ul style="list-style-type: none"> • Cells did not lyse completely • Experiment was not performed at optimal time after apoptosis induction • Plate read at incorrect wavelength • Old DTT used | <ul style="list-style-type: none"> • Resuspend the cell pellet in the lysis buffer and incubate as described in the datasheet • Perform a time-course induction experiment for apoptosis • Check the wavelength listed in the datasheet and the filter settings of the instrument • Always use freshly thawed DTT in the cell lysis buffer |
| High Background | <ul style="list-style-type: none"> • Increased amount of cell lysate used • Increased amounts of components added due to incorrect pipetting • Incubation of cell samples for extended periods • Use of expired kit or improperly stored reagents • Contaminated cells | <ul style="list-style-type: none"> • Refer to datasheet and use the suggested cell number to prepare lysates • Use calibrated pipettes • Refer to datasheet and incubate for exact times • Always check the expiry date and store the individual components appropriately • Check for bacteria/ yeast/ mycoplasma contamination |
| Lower signal levels | <ul style="list-style-type: none"> • Cells did not initiate apoptosis • Very few cells used for analysis • Use of samples stored for a long time • Incorrect setting of the equipment used to read samples • Allowing the reagents to sit for extended times on ice | <ul style="list-style-type: none"> • Determine the time-point for initiation of apoptosis after induction (time-course experiment) • Refer to datasheet for appropriate cell number • Use fresh samples or aliquot and store and use within one month for the assay • Refer to datasheet and use the recommended filter setting • Always thaw and prepare fresh reaction mix before use |
| Samples with erratic readings | <ul style="list-style-type: none"> • Uneven number of cells seeded in the wells • Samples prepared in a different buffer • Adherent cells dislodged and lost at the time of experiment • Cell/ tissue samples were not completely homogenized • Samples used after multiple freeze-thaw cycles • Presence of interfering substance in the sample • Use of old or inappropriately stored samples | <ul style="list-style-type: none"> • Seed only equal number of healthy cells (correct passage number) • Use the cell lysis buffer provided in the kit • Perform experiment gently and in duplicates/triplicates; apoptotic cells may become floaters • Use Dounce homogenizer (increase the number of strokes); observe efficiency of lysis under microscope • Aliquot and freeze samples, if needed to use multiple times • Troubleshoot as needed • Use fresh samples or store at correct temperatures until use |
| Unanticipated results | <ul style="list-style-type: none"> • Measured at incorrect wavelength • Cell samples contain interfering substances | <ul style="list-style-type: none"> • Check the equipment and the filter setting • Troubleshoot if it interferes with the kit (run proper controls) |
| General issues | <ul style="list-style-type: none"> • Improperly thawed components • Incorrect incubation times or temperatures • Incorrect volumes used • Air bubbles formed in the well/tube | <ul style="list-style-type: none"> • Thaw all components completely and mix gently before use • Refer to datasheet & verify the correct incubation times and temperatures • Use calibrated pipettes and aliquot correctly |

| | | |
|--|---|---|
| | <ul style="list-style-type: none"> • Substituting reagents from older kits/ lots • Use of a different 96-well plate | <ul style="list-style-type: none"> • Pipette gently against the wall of the well/tubes • Use fresh components from the same kit • Fluorescence: Black plates; Absorbance: Clear plates |
| <p>Note# The most probable cause is listed under each section. Causes may overlap with other sections.</p> | | |

For research use only! Not to be used in humans.

Our promise

If the product does not perform as described on this datasheet, we will offer a refund or replacement. For more details, please visit <http://www.apexbt.com/> or contact our technical team.

Tel: +1-(832)696-8203

Fax: +1-832-641-3177

Email: sales@apexbt.com