

High Performance Liquid Chromatography

Analysis Information

Product Name: S6 Kinase Substrate Peptide 32
Operator: David
Injection Date: 2/16/2019 9:24:42 PM
Batch No.: 1


HPLC Condition

Column : ZORBAX Eclipse Pluse C18 3.5 μ m 4.6mmx100mm
Solvent A: H2O+0.1%TFA
Solvent B: ACN+0.1%TFA
Gradient:

Time	A	B
0.0min	95%	5%
2.0min	95%	5%
7.0min	5%	95%
10.00min	0%	100%
12.00min	0%	100%
12.10min	95%	5%
15.00min	Stop	Stop

Flow rate: 1mL/min
Wavelength: 220nm
Volume: 10 μ l

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	5.069	1.69172e4	98.2832
2	5.421	99.36804	0.5773
3	5.774	54.85356	0.3187
4	5.822	141.28375	0.8208

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com