

High Performance Liquid Chromatography

1. Analysis Information

Product Name:	Z-VAD-FMK
Operator:	David
Injection Date:	2016-7-26 13:47:54
Batch No.:	14


2. HPLC Condition

Column:	Athena C18, 3 μ m, 2.1mm \times 100 mm
Solvent A:	0.1% H ₃ PO ₄ in 100% Acetonitrile
Solvent B:	0.1% H ₃ PO ₄ in 100% Water

Gradient:			
	Time	A	B
	0.0min	0%	100%
	2.0min	0%	100%
	6.0min	95%	5%
	9.0min	95%	5%
	10.0min	0%	100%
	10.0min	Stop	Stop

Flow rate:	0.4ml/min
Wavelength:	215nm
Volume:	10 μ l

3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	2.984	6370.46484	99.5970
2	4.078	8.93393	0.1397
3	5.164	16.84189	0.2633