

High Performance Liquid Chromatography

Analysis Information

Product Name: AEBSF.HCl

Operator: David

Injection Date: 12/6/2021 1:37:11 PM

Batch No.: 19

HPLC Condition

Column : Athena C18, 3 μ m, 2.1mm $\times$ 100 mm

Solvent A: ACN+0.1%TFA

Solvent B: H2O+0.1%TFA


Gradient:

Time	A	B
0.0min	0%	100%
1.5min	0%	100%
6.0min	95%	5%
9.0min	95%	5%
9.0min	Stop	Stop

Flow rate: 0.46ml/min

Wavelength: 230nm

Volume: 3 μ l



Rank	RetTime(min)	Area(mAU*s)	Area %
1	5.011	1.05844e4	98.2905
2	5.688	8.61966	0.0800
3	5.979	10.19715	0.0947
4	6.577	12.45232	0.1156
5	6.698	133.59261	1.2406
6	7.999	19.22607	0.1785

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com

