

High Performance Liquid Chromatography

Analysis Information

Product Name: PMSF

Operator: David

Injection Date: 3/1/2023 11:54:01 PM

Batch No.: 10

HPLC Condition

Column : Athena C18, 3 μ m, 2.1mm $\times$ 100 mm

Solvent A: ACN+0.1%TFA

Solvent B: H2O+0.1%TFA


Gradient:

Time	A	B
0.0min	0%	100%
1.5min	0%	100%
6.0min	95%	5%
9.0min	95%	5%
9.0min	Stop	Stop

Flow rate: 0.46ml/min

Wavelength: 210nm

Volume: 3 μ l



Rank	RetTime(min)	Area(mAU*s)	Area %
1	5.622	24.33952	0.6487
2	6.544	3715.76416	99.0327
3	7.098	2.56303	0.0683
4	7.438	4.21649	0.1124
5	7.746	5.17412	0.1379

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com

