

High Performance Liquid Chromatography

Analysis Information

Product Name: LCL161

Operator: David

Injection Date: 4/21/2020 4:53:31 PM

Batch No.: 2

HPLC Condition

Column : ZORBAX Eclipse Pluse C18 3.5 μ m 4.6mmx100mm

Solvent A: H2O+0.1%TFA

Solvent B: ACN+0.1%TFA

Gradient:

Time	A	B
0.0min	80%	20%
2.0min	80%	20%
7.0min	5%	95%
10.00min	0%	100%
12.00min	0%	100%
15.00min	Stop	Stop


Flow rate: 1mL/min

Wavelength: 280nm

Volume:

5µl

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	6.031	25.05817	0.2356
2	6.163	1.05563e4	99.2413
3	7.536	17.49101	0.1644
4	7.852	38.15308	0.3587

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com