

High Performance Liquid Chromatography

Analysis Information

Product Name: ONO-AE3-208

Operator: David

Injection Date: 3/24/2021 1:00:59 PM

Batch No.: 2

HPLC Condition

Column : ZORBAX Eclipse Pluse C18 3.5 μ m 4.6mmx100mm

Solvent A: H2O+0.1%TFA

Solvent B: ACN+0.1%TFA

Gradient:


Time	A	B
0.0min	80%	20%
2.0min	80%	20%
16.0min	20%	80%
17.0min	0%	100%
20.0min	0%	100%
23.0min	Stop	Stop

Flow rate: 1ml/min

Wavelength: 280nm

Volume: 5 μ l

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	11.940	26.71372	0.5821
2	12.420	4485.29492	97.7336
3	14.030	77.29967	1.6843

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com