

High Performance Liquid Chromatography

Analysis Information

Product Name: QNZ (EVP4593)

Operator: David

Injection Date: 4/22/2020 1:33:51 PM

Batch No.: 2

HPLC Condition

Column : ZORBAX Eclipse Pluse C18 3.5 μ m 4.6mmx100mm

Solvent A: H2O+0.1%TFA

Solvent B: ACN+0.1%TFA

Gradient:


Time	A	B
0.0min	80%	20%
2.0min	80%	20%
7.0min	5%	95%
10.00min	0%	100%
12.00min	0%	100%
15.00min	Stop	Stop

Flow rate: 1mL/min

Wavelength: 280nm

Volume: 5 μ l

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	5.424	2.30720	0.0540
2	5.665	2.11856	0.0496
3	6.111	3.04179	0.0712
4	6.362	4245.01514	99.4180
5	6.728	2.86131	0.0670
6	6.871	4.29319	0.1005
7	7.010	7.39429	0.1732
8	7.431	2.83609	0.0664

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com