

High Performance Liquid Chromatography

Analysis Information

Product Name: Bz-Arg-OEt.HCl

Operator: David

Injection Date: 11/15/2021 6:56:28 PM

Batch No.: 2

HPLC Condition

Column : Athena C18, 3 μ m, 2.1mm $\times$ 100 mm

Solvent A: ACN+0.1%TFA

Solvent B: H2O+0.1%TFA


Gradient:

Time	A	B
0.0min	0%	100%
2.0min	0%	100%
7.0min	95%	5%
10.0min	95%	5%
10.0min	Stop	Stop

Flow rate: 0.45ml/min

Wavelength: 230nm

Volume: 3 μ l



Rank	RetTime(min)	Area(mAU*s)	Area %
1	5.584	57.77309	0.7602
2	5.657	39.43085	0.5189
3	6.050	7492.63721	98.5945
4	6.432	9.60396	0.1264

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com