

High Performance Liquid Chromatography

1. Analysis Information

Product Name: HATU
 Operator: David
 Injection Date: 8/15/2013 5:42:47 PM
 Batch No.: 1


2. HPLC Condition

Column: Waters XBridge C18, 3 μ m, 3 mm $\times$ 100 mm
 Solvent A: 0.1% H₃PO₄ in 100% Acetonitrile
 Solvent B: 0.1% H₃PO₄ in 100% Water

Gradient:	Time	A	B
	0.0min	0%	100%
	4.0min	0%	100%
	6.0min	95%	5%
	10.0min	95%	5%
	10.0min	Stop	Stop

Flow rate: 0.5ml/min
 Wavelength: 254nm
 Volume: 10 μ l

3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	1.308	4005.22314	99.0146
2	4.160	39.86214	0.9854