

High Performance Liquid Chromatography


1. Analysis Information

Product Name:	EPZ-6438
Operator:	David
Injection Date:	1/21/2014 6:47:29 PM
Batch No.:	1

2. HPLC Condition

Column:	Athena C18 (3 μ m, 2.1 mm \times 100 mm)																									
Solvent A:	0.1% H3PO4 in 100% Acetonitrile																									
Solvent B:	0.1% H3PO4 in 100% Water																									
Gradient:	<table border="1"> <thead> <tr> <th>Time</th> <th>A</th> <th>B</th> </tr> </thead> <tbody> <tr> <td>0.0min</td> <td>10%</td> <td>90%</td> </tr> <tr> <td>2.0min</td> <td>65%</td> <td>35%</td> </tr> <tr> <td>3.5min</td> <td>95%</td> <td>5%</td> </tr> <tr> <td>4.5min</td> <td>10%</td> <td>90%</td> </tr> <tr> <td>5.5min</td> <td>10%</td> <td>90%</td> </tr> <tr> <td>10.0min</td> <td>10%</td> <td>90%</td> </tr> <tr> <td>10.0min</td> <td>Stop</td> <td>Stop</td> </tr> </tbody> </table>		Time	A	B	0.0min	10%	90%	2.0min	65%	35%	3.5min	95%	5%	4.5min	10%	90%	5.5min	10%	90%	10.0min	10%	90%	10.0min	Stop	Stop
Time	A	B																								
0.0min	10%	90%																								
2.0min	65%	35%																								
3.5min	95%	5%																								
4.5min	10%	90%																								
5.5min	10%	90%																								
10.0min	10%	90%																								
10.0min	Stop	Stop																								
Flow rate:	0.4ml/min																									
Wavelength:	254nm																									
Volume:	10 μ l																									

3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	2.879	8270.89551	99.6557
2	3.172	18.73387	0.2257
3	5.361	9.84123	0.1186