



# High Performance Liquid Chromatography

## 1. Analysis Information

Product Name: CA-074 Me  
Operator: David  
Injection Date: 11/22/2014 5:39:53 PM  
Batch No.: 4

## 2. HPLC Condition


Column: Athena C18, 3  $\mu$ m, 2.1mm  $\times$  100 mm  
Solvent A: 0.1% H<sub>3</sub>PO<sub>4</sub> in 100% Acetonitrile  
Solvent B: 0.1% H<sub>3</sub>PO<sub>4</sub> in 100% Water

Gradient:

Time	A	B
0.0min	10%	90%
3.5min	95%	5%
8.0min	95%	5%
9.0min	10%	90%
10.0min	10%	90%
10.0min	Stop	Stop

Flow rate: 0.4ml/min  
Wavelength: 220nm  
Volume: 10 $\mu$ l

## 3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	3.605	86.87156	1.5043
2	3.873	5688.14404	98.4957