

High Performance Liquid Chromatography

Analysis Information

Product Name: ITE

Operator: David

Injection Date: 11/1/2021 12:51:14 PM

Batch No.: 3

HPLC Condition

Column : Athena C18, 3 μ m, 2.1mm $\times$ 100 mm

Solvent A: ACN+0.1%TFA

Solvent B: H2O+0.1%TFA

Gradient:


Time	A	B
0.0min	10%	90%
5.0min	95%	5%
9.0min	95%	5%
9.0min	Stop	Stop

Flow rate: 0.46ml/min

Wavelength: 210nm

Volume: 3 μ l

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	4.506	74.76255	2.5404
2	4.790	14.11973	0.4798
3	5.360	2854.09521	96.9799

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com