

High Performance Liquid Chromatography

Analysis Information

Product Name: C598-0466

Operator: David

Injection Date: 7/11/2019 4:38:15 PM

Batch No.: 2

HPLC Condition

Column : ZORBAX Eclipse Pluse C18 3.5 μ m 4.6mmx100mm

Solvent A: H2O+1‰ TFA

Solvent B: ACN+1‰ TFA

Gradient:


Time	A	B
0.0min	80%	20%
2.0min	80%	20%
7.0min	5%	95%
10.0min	0%	100%
12.0min	0%	100%
12.1min	80%	20%
15.0min	Stop	Stop

Flow rate: 1ml/min

Wavelength: 214nm

Volume: 5 μ l

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	5.412	8883.15137	98.3767
2	5.629	32.58888	0.3609
3	5.696	12.56494	0.1392
4	5.990	57.03084	0.6316
5	6.918	44.39239	0.4916

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com