

High Performance Liquid Chromatography

1. Analysis Information

Product Name: ICP 103
 Operator: David
 Injection Date: 5/18/2014 21:13:21 PM
 Batch No.: 1


2. HPLC Condition

Column: xBridge C18, 4.6*150mm, 3.5 μ m
 Solvent A: 0.1%Trifluoroacetic in 100% Acetonitrile
 Solvent B: 0.1%Trifluoroacetic in 100% Water

Gradient:	Time	A	B
	0.0min	50%	50%
	15.0min	90%	10%

Flow rate: 0.6ml/min
 Wavelength: 360nm
 Volume: 10 μ l

3. Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	8.703	84038	0.55
2	10.656	292071	1.93
3	11.248	14686910	96.99
4	13.495	79789	0.53