

High Performance Liquid Chromatography

Analysis Information

Product Name: AM251

Operator: David

Injection Date: 6/29/2023 15:22:33

Batch No.: 5

HPLC Condition

Column : ZORBAX Eclipse Pluse C18 3.5 μ m 4.6mmx100mm

Solvent A: H2O+0.1%TFA

Solvent B: ACN+0.1%TFA

Gradient:


Time	A	B
0.0min	80%	20%
2.0min	80%	20%
16.0min	20%	80%
17.0min	0%	100%
20.0min	0%	100%
23.0min	Stop	Stop

Flow rate: 1ml/min

Wavelength: 265nm

Volume: 2 μ l

Result



Rank	RetTime(min)	Area(mAU*s)	Area %
1	14.957	19.80	0.62
2	15.318	3142.77	98.64
3	19.065	23.38	0.73

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com