

High Performance Liquid Chromatography

Analysis Information

Product Name: Suplatast Tosylate

Operator: David

Injection Date: 12/17/2021 1:10:53 PM

Batch No.: 2

HPLC Condition

Column : Athena C18, 3 μ m, 2.1mm $\times$ 100 mm

Solvent A: ACN+0.1%TFA

Solvent B: H2O+0.1%TFA


Gradient:

Time	A	B
0.0min	0%	100%
1.5min	0%	100%
6.0min	95%	5%
9.0min	95%	5%
9.0min	Stop	Stop

Flow rate: 0.46ml/min

Wavelength: 254nm

Volume: 3 μ l



Rank	RetTime(min)	Area(mAU*s)	Area %
1	5.081	85.60449	98.1130
2	5.419	22.37241	1.0525
3	5.525	7.22294	0.3398
4	5.605	10.51693	0.4947

Caution**FOR RESEARCH PURPOSES ONLY.****NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.****APExBIO Technology****www.apexbt.com**

7505 Fannin street, Suite 410, Houston, TX 77054.

