

High Performance Liquid Chromatography

Analysis Information

Product Name: HX 531

Operator: David

Injection Date: 02/25/2025 15:28:41

Batch No.: 2

HPLC Condition

Column : ZORBAX Eclipse Pluse C18 3.5 μ m 4.6mmx100mm

Solvent A: ACN+0.1%TFA

Solvent B: H2O+0.1%TFA


Gradient:

Time	A	B
0.0min	75%	25%
2.0min	75%	25%
5.0min	100%	0%
12.0min	100%	0%
15.0min	Stop	Stop

Flow rate: 1ml/min

Wavelength: 265nm

Volume: 3 μ l



Rank	RetTime(min)	Area(mAU*s)	Area %
1	4.118	27.49	0.81
2	5.536	16.69	0.49
3	6.326	3246.73	95.23
4	6.599	88.13	2.58
5	7.002	11.53	0.34
6	7.610	18.79	0.55

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

