

# High Performance Liquid Chromatography

## Analysis Information


**Product Name:** SKF38393 HCl  
**Operator:** David  
**Injection Date:** 1/4/2019 10:26:40 AM  
**Batch No.:** 3

## HPLC Condition

**Column :** ZORBAX Eclipse Pluse C18 3.5 $\mu$ m 4.6mmx100mm  
**Solvent A:** H2O+1‰ TFA  
**Solvent B:** ACN+1‰ TFA  
**Gradient:**

| Time    | A    | B    |
|---------|------|------|
| 0.0min  | 100% | 0%   |
| 2.0min  | 100% | 0%   |
| 7.0min  | 5%   | 95%  |
| 10.0min | 5%   | 95%  |
| 10.0min | Stop | Stop |

**Flow rate:** 1ml/min  
**Wavelength:** 270nm  
**Volume:** 5 $\mu$ l



| Rank | RetTime(min) | Area(mAU*s) | Area %  |
|------|--------------|-------------|---------|
| 1    | 4.721        | 8.76676     | 0.7793  |
| 2    | 5.784        | 1116.15344  | 99.2207 |

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com