

Mass Spectrometry

Analysis Information

Product Name:	Sodium citrate
Operator:	David
Injection Date:	8/13/2019 5:19:08 PM
Inj. Volume:	1
Batch No.:	2


Acquisition Parameter

Probe:	ESI
Gas Temp:	350°C
Gas Flow:	11L/min
Nebulizer:	50psi
VCap:	4000v
Fragmentor:	70v
Reference:	m/z 121.050873, 922.009798
T.Flow:	0.3ml/min
B.conc:	20%H ₂ O/80%ACN

Result

20190813-68-04-2 9 (0.156)

2: Scan ES-
2.03e6



Molecular Weight: 258.07

m/z([M-3Na]-) Base Peak: 191.22

m/z([M+Na]+) Base Peak: 281.22

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com