

Mass Spectrometry

Analysis Information

Product Name: Golgi-Tracker Green

Operator: David

Injection Date: 09/30/2024 9:11:39 PM

Inj. Volume: 1

Batch No.: 2

Acquisition Parameter

Probe: ESI

Gas Temp: 350°C

Gas Flow: 11L/min

Nebulizer: 50psi

VCap: 4000v


Fragmentor: 70v

Reference: m/z 121.050873, 922.009798

T.Flow: 0.3ml/min

B.conc: 20%H₂O/80%ACN

Result



Molecular Weight:

601.62

m/z([M-H]-) Base Peak:

600.6

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX-BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

