

High Performance Liquid Chromatography

Analysis Information

Product Name: NVS-STG2

Operator: David

Injection Date: 02/14/2025 13:30:06

Batch No.: 2

HPLC Condition

Column : ZORBAX Eclipse Pluse C18 3.5 μ m 4.6mmx100mm

Solvent A: ACN+0.1%TFA

Solvent B: H2O+0.1%TFA


Gradient:

Time	A	B
0.0min	75%	25%
2.0min	75%	25%
5.0min	100%	0%
12.0min	100%	0%
15.0min	Stop	Stop

Flow rate: 1ml/min

Wavelength: 280nm

Volume: 5 μ l



Rank	RetTime(min)	Area(mAU*s)	Area %
1	5.966	1640.14	99.04
2	7.492	15.93	0.96

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com