

Mass Spectrometry

Analysis Information

Product Name: L-Cysteine HCl

Operator: David

Injection Date: 07/15/2020

Inj. Volume: 1

Batch No.: 2

Acquisition Parameter

Probe: ESI

Gas Temp: 350°C

Gas Flow: 11L/min

Nebulizer: 50psi

VCap: 4000v

Fragmentor: 70v

Reference: m/z 121.050873, 922.009798

T.Flow: 0.3ml/min


B.conc: 20%H₂O/80%ACN

Result

NT-S20200715-010

20200715-C6479-2 11 (0.182)

1: Scan ES+
7.80e7



Molecular Weight: 157.62

m/z([M-Cl]-)Base Peak: 122.06

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com