

# High Performance Liquid Chromatography

## Analysis Information


**Product Name:** ABTS diammonium salt  
**Operator:** David  
**Injection Date:** 6/13/2022 7:00:45 PM  
**Batch No.:** 3

## HPLC Condition

**Column :** Athena C18, 3  $\mu$ m, 2.1mm  $\times$  100 mm  
**Solvent A:** ACN+0.1%TFA  
**Solvent B:** H2O+0.1%TFA  
**Gradient:**

Time	A	B
0.0min	0%	100%
2.0min	0%	100%
7.0min	95%	5%
10.0min	95%	5%
10.0min	Stop	Stop

**Flow rate:** 0.45ml/min  
**Wavelength:** 230nm  
**Volume:** 3 $\mu$ l



Rank	RetTime(min)	Area(mAU*s)	Area %
1	6.070	19.18747	0.0872
2	6.206	3.34924	0.0152
3	6.399	2.19694e4	99.8026
4	7.972	17.58243	0.0799
5	8.821	3.33341	0.0151

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: [info@apexbt.com](mailto:info@apexbt.com)

