

### **Product Information**

# Fluorescein Active Caspase-9 Staining Kit

#### I. Kit Contents:

Component	K2051-25	K2051-100	Part Number
	25 assays	100 assays	
FITC-LEHD-FMK	25 μl	100 μ1	K2051-C-1
Wash Buffer	50 ml	2 x 100 ml	K2051-C-2
Z-VAD-FMK	10 μ1	10 μ1	A1902

#### II. Introduction:

Caspases activation is curial in apoptosis. Using Fluorescein Caspase-9 Staining Kit is an easy and sensitive way of detecting activated caspase-3 in living cells. This assay uses the Caspase-9 inhibitor, LEHD-FMK, coupled to FITC (FITC-LEHD-FMK) as a marker. In apoptotic cells, the cell permeable and nontoxic FITC-LEHD-FMK irreversibly binds to activated Caspase-9.

### III. Caspase-9 Assay Protocol:

A. Staining Procedure:

- 1. Induce apoptosis in cells (1 x  $10^6$ /ml) by desired method. Concurrently incubate a control culture without induction. An additional negative control can be prepared by adding the caspase inhibitor Z-VAD-FMK at 1  $\mu$ l/ml to an induced culture to inhibit caspase activation.
- 2. Aliquot 300 µl each of the induced and control cultures into eppendorf tubes.
- 3. Add 1 µl of FITC-LEHD-FMK into each tube and incubate for 0.5 1 hour at 37°C incubator with 5 % CO<sub>2</sub>.
- 4. Centrifuge cells at 3000 rpm for 5 minutes and remove supernatant.
- 5. Resuspend cells in 0.5 ml of Wash Buffer, and centrifuge again.
- 6. Repeat Step 5.

Proceed to B, C, or D depending on methods of analysis.

B. Quantification by Flow Cytometry:

For flow cytometric analysis, resuspend cells in 300 µl of Wash buffer. Put samples on ice. Analyzing samples by flow cytometry using the FL-1 channel.

C. Detection by Fluorescence Microscopy:

For fluorescence microscopic analysis, resuspend cells in 100 µl Wash buffer. Add one drop of the cell suspension onto a microslide and cover with a coverslip. Observe cells under a fluorescence microscope using FITC filter. Caspase positive cells appear to have brighter green signals, whereas caspase negative control cells show much weaker signal.

D. Analysis by Fluorescence Plate Reader:

For analysis with fluorescence plate reader, resuspend cells in 100  $\mu$ l Wash Buffer and then transfer the cell suspension to each well of the black microtiter plate. Measure the fluorescence intensity at Ex/Em = 485/535 nm. For control, use wells containing unlabeled cells.



## **General Troubleshooting Guide:**

Cell density is higher than recommended     Cells were not washed well with wash buffer after staining     Cells were Incubated for extended period of time     Use of extremely confluent cells     Contaminated cells	<ul> <li>Refer to data sheet and use the suggested cell number</li> <li>Use the wash buffer provided, and as instructed in the datasheet</li> <li>Refer to data sheets and incubate for exact times</li> <li>Perform assay when cells are at 70-95% confluency</li> <li>Check for bacteria/ yeast/ mycoplasma contamination</li> </ul>	
Cells were Incubated for extended period of time     Use of extremely confluent cells     Contaminated cells	datasheet • Refer to data sheets and incubate for exact times • Perform assay when cells are at 70-95% confluency	
Use of extremely confluent cells     Contaminated cells	<ul> <li>Refer to data sheets and incubate for exact times</li> <li>Perform assay when cells are at 70-95% confluency</li> </ul>	
Contaminated cells	• Perform assay when cells are at 70-95% confluency	
	Check for bacteria/ yeast/ mycoplasma contamination	
Cells did not initiate apoptosis	• Determine the time-point for initiation of apoptosis after	
Very few cells used for analysis	induction (time-course experiment)	
Incorrect setting of the equipment used to read samples	• Refer to data sheet for appropriate cell number	
Use of expired kit or improperly stored reagents	• Refer to data sheet and use the recommended filter setting	
	• Always check the expiry date and store the components	
	appropriately	
Old (unhealthy) cells used	• Seed healthy cells and make sure cells are healthy prior to	
Adherent cells were dislodged and washed away prior to	induction of apoptosis	
assaying	• Collect all cells (both attached and dislodged) after	
Incorrect incubation times or temperatures	induction for accurate results	
Incorrect volumes used	• Refer to datasheet & verify correct incubation times and	
	temperatures	
	• Use calibrated pipettes and aliquot correctly	
• • • a	Incorrect setting of the equipment used to read samples Use of expired kit or improperly stored reagents  Old (unhealthy) cells used Adherent cells were dislodged and washed away prior to assaying Incorrect incubation times or temperatures	

For research use only! Not to be used in humans.

# Our promise

If the product does not perform as described on this datasheet, we will offer a refund or replacement. For more details, please visit <a href="http://www.apexbt.com/">http://www.apexbt.com/</a> or contact our technical team.

Tel: +1-(832)696-8203 Fax: +1-832-641-3177 Email: sales@apexbt.com

Tel: +1-832-696-8203; Fax: +1-832-641-3177 http://www.apexbt.com/; Email: sales@apexbt.com.