

Mass Spectrometry

Analysis Information

Product Name: Anemoside A3

Operator: David

Injection Date: 3/7/2022 8:09:39 PM

Inj. Volume: 1

Batch No.: 3

Acquisition Parameter

Probe: ESI

Gas Temp: 350°C

Gas Flow: 11L/min

Nebulizer: 50psi

VCap: 4000v

Fragmentor: 70v

Reference: m/z 121.050873, 922.009798


T.Flow: 0.3ml/min

B.conc: 20%H₂O/80%ACN

Result

220307 129724-84-1 107 (3.939) Cm (107:111)

1: Scan ES+
1.44e6



Molecular Weight: 750.96

m/z([M+Na]+) Base Peak: 773.8

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com