Certificate of Analysis

Product Name:	HyperTrap Butyl 4FF Column
Cat. No.:	PC1020
Batch No.:	26001
Production Date:	Nov 10 2026
Expiration Date:	Nov 09 2031
Storage Condition:	Store at 4°C

Physical and Chemical Properties

Item	Standard	Result
Appearance:	Milky white, translucent spherical particles; no visible impurities	Pass
Mean Particle Size	80 - 100 μm	89.8 μm
Dynamic Binding Capacity	15 - 25 mg BSA/mL	20.7 mg BSA/mL
Bioburden	≤100 CFU/mL	Pass

Analytical Data

THE COMPONENTS OF THIS PRODUCT HAVE BEEN TESTED AND MEET ALL QUALITY SPECIFICATIONS

Signed by: Aixuan Li

Quality Assurance

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APExBIO Technology
www.apexbt.com
7505 Fannin street, Suite 410, Houston, TX 77054.
Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com